

EC Declaration of Conformity

We, Intersurgical Ltd (address: Crane House, Molly Millars Lane, Wokingham, Berkshire, RG41 2RZ, United Kingdom) as manufacturer, hereby declare under the sole responsibility of that the below mentioned devices comply with European Medical Devices Directive 93/42/EEC with all amendments and transposing legislation.

Authorised Representative in the European Economic Area (EEA):
UAB Intersurgical (address: Arnioniu g. 60, Pabradė, LT-18170, Lithuania)

Filters

These are class IIA devices in accordance with rule 2 of Annex IX of the Medical Devices Directive 93/42/EEC (classification.doc)

GMDN code – 60837, 60839, 61097, 60506

Essential requirements checklist is on IQR139.

Internally manufactured components master data is on EFACS (Parts Master) and design drawings are on IQR69 (Product and Mould Drawing List).

Internally manufactured finished products master data, production drawings and Master Product Formulae are on EFACS (Master Product Formula, Parts Master) and IQR69 (Product and Mould Drawing List).

Externally manufactured materials, components, design drawings and products are on EFACS (Parts Master), IQR98 (Incoming Product Specification) and IQR69 (Product and Mould Drawing List).

Instructions for use, pack inserts and labels are as detailed on EFACS (Master Product Formula) and IQR107 (Index of Controlled Artworks).

Label content is on EFACS (Master Product Formula) and Labels Printing System.

Product realisation processes are referenced in IQM section 7.1

Product Codes

As listed in EFACS MPF Details under group DCFILTER.DOC.

This range is subject to the procedure set out in Annex 2 excluding section 4 of Directive 93/42/EEC as amended by 2007/47/EC, under the supervision of Notified Body Number 1639, SGS Belgium NV, SGS House Noorderlaan 87-2030, Antwerpen, Belgium

EC Certificate Full Quality Assurance System GB19/964232 has been issued for the management system of Intersurgical Ltd, which has been assessed and certified according to the requirements of Annex 2 excluding section 4 of Directive 93/42/EEC as amended by 2007/47/EC.



Ivan Seniut
Group Quality and Regulatory Affairs Director
Duly authorised for and on behalf of Intersurgical Ltd
Crane House, Molly Millars Lane, Wokingham,
Berkshire, RG41 2RZ, United Kingdom

Issue 19
Valid from 1 January 2021
DCFILTER.DOC